

Specifications

Features	Specifications
Methodology	Reflectance Photometry
Detection	Photosensitive Diode
Calibration	Automatic
Throughput	Single Test Option: 60 tests/hour, continuous test option: 120 test/hour
Memory	Last 2000 results
Strip Incubation Time	1 min
Wavelength of Monochromatic LED	525 nm and 635 nm
Strips Available	1-14 Parameters (108 mm x 5 mm)
Strip Combinations per Analyzer	Over 15 combinations
Data Entry Capabilities	Operator/Patient ID-Manual Entry and Barcode Reader (Up to 20 characters)
Analyzer Ports	Standard RS232C Port for Barcode Reader, Data transfer or software update, USB Port for Data Transfer or software update, 25 Pin Parallel Port for External Printer
Capabilities	Internal Thermal Printer (included), Optional External Printer (not included), RS232C Barcode Reader (optional), USB or RS232C Data Transfer Cable (optional)
Major Readable Barcodes	Code 128, Code 39, Codabar (NW-7), EAN 8, EAN 13, Interleave 25, UPCA, UPCE
Available Languages on Screen	English and additional languages
Analyzer Operating Conditions	0-40 °C (32-104 °F); ≤85% Relative Humidity (non-condensing)
Strip Operating Conditions	15-30 °C (59-86 °F); ≤85% Relative Humidity (non-condensing)
Storage Conditions	-5- 50 °C (23-122 °F); 90% Relative Humidity
Power Source	100-240 V AC, 50/60 Hz, 35 VA
Weight	1.85 kg (4.1 lbs)
Dimensions (L x W x H)	27.5 cm x 18.3 cm x 13.9 cm (10.8" x 7.2" x 5.5")
Display Dimensions (L x W)	10.8 cm x 6.5 cm (4.2" x 2.6")

CE marked for sale in the European Community

Ordering Information

Product Name	Catalog No.	Components	Kit Box Dimensions (L X W X H) & Weight	Carton Dimensions (L X W X H) & Weight	Number of Kits/Carton	
<i>Mission® U120 Smart Urine Analyzer</i>	U117-101	1 Urine Analyzer	1 Power Cord	49 cm x 28 cm x 22.5 cm, 3.4 kg	1	
		1 Quick Start Guide	1 Instruction Manual	19.3" x 11.0" x 8.9", 119.9 oz		
<i>Mission® U120 Smart Urine Analyzer (with Barcode Reader)</i>	U117-111	1 Urine Analyzer	1 Power Cord	49 cm x 28 cm x 22.5 cm, 3.8 kg	1	
		1 Quick Start Guide	1 Instruction Manual	19.3" x 11.0" x 8.9", 134.0 oz		
<i>Mission® Barcode Reader</i>	U221-111	1 Barcode Reader (RS232C)	1 Serial Splitter Cable (RS232C)	23.6 cm x 10.8 cm x 7.8 cm; 0.482 kg	27.5 cm x 24 cm x 25 cm, 2.46 kg	4
		9.3" x 4.3" x 3.1"; 17.0 oz	10.8" x 9.4" x 9.8"; 86.8 oz			
<i>Mission® Printer Paper Rolls</i>	U121-101	4 Printer Paper Rolls	Thermal Paper (0.06 m x 20 m): 200 results/roll	12.0 cm x 12.0 cm x 6.5 cm; 0.36 kg	63.0 cm x 37.0 cm x 30.0 cm; 19.4 kg	50
			4.7" x 4.7" x 2.6"; 12.7 oz	24.8" x 14.6" x 11.8"; 684.3 oz		
		Sticker Paper (0.06 m x 9 m): 100 results/roll	12.0 cm x 12.0 cm x 6.5 cm; 0.4 kg	63.0 cm x 37.0 cm x 30.0 cm; 21.4 kg		
			4.7" x 4.7" x 2.6"; 14.1 oz	24.8" x 14.6" x 11.8"; 754.9 oz		
<i>Mission® U120 Smart Data Transfer Kit</i>	U127-131	1 Data Transfer Cable (RS232C or USB)	1 Package Insert	16.0 cm x 13.0 cm x 3.5 cm; 0.147 kg	25.0 cm x 21.0 cm x 15.0 cm; 1.36 kg	8
		6.3" x 5.1" x 1.4"; 5.2 oz	9.8" x 8.3" x 5.9"; 48.0 oz			

We also offer other rapid diagnostic and medical products for:

Blood Glucose, Lateral flow, POCT Products, Immunoassay, Molecular Diagnostics and more.

Contact us for worldwide distribution and custom manufacturing (OEM) opportunities



aconlabs.com

ACON Laboratories, Inc.
10125 Mesa Rim Road, San Diego, CA 92121, U.S.A.
Tel: 1.858.875.8000
Fax: 1.858.200.0729
Email: info@aconlabs.com

Attention: By approving the enclosed design draft, you (ACON's Customer) accept all responsibility for the accuracy of the design. If an error is detected following the printing or manufacturing of a material, you (ACON's Customer) are responsible for the cost of any inventory which is deemed unsuitable for sale and/or not compliant with local regulations .

US

OUS

DOMESTIC

OTHER

Description

Mission U120 Smart Sell Sheet

Part Number

1150955001

Size

210x280mm

Printing Contents

L Number

Size

Designer

rene

Design Date/Version

Nov 8, 2017 /D

Artwork checked by

Material

230g 双铜 +油性上光

Checked by

Approved by Customer

Approved by Marketing/Sales

Approved by P.M.T.

Approved by QA

Effective Date